



കേരള സർക്കാർ
Government of Kerala
2014



Regn. No. KERBIL/2012/45073
dated 5-9-2012 with RNI

Reg. No. KL/TV(N)/634/2012-14

tl cf Kk ddv KERALA GAZETTE

ആധികാരികമായി പ്രസിദ്ധപ്പെടുത്തുന്നത്
PUBLISHED BY AUTHORITY

വാല്യം 3 Vol. III	} തിരുവനന്തപുരം, ചൊവ്വ Thiruvananthapuram, Tuesday	2014 ജനുവരി 21 21st January 2014	} നമ്പർ No.	} 3
		1189 മകരം 8 8th Makaram 1189		
		1935 മാഘം 1 1st Magha 1935		

PART III

Stores Purchase

Health Services Department

ദർഘാസ് പരസ്യം

നമ്പർ എം.എസ്.പി.1-79723/2013/ആ.വ.ഡ.

2013 ഡിസംബർ 28.

ആരോഗ്യവകുപ്പിന്റെ കീഴിലുള്ള സ്ഥാപനങ്ങളിൽ താഴെ പറയുന്ന സാധനങ്ങൾ വിതരണം ചെയ്യുന്നതിന് മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

ദർഘാസ് നമ്പരും ഉപകരണത്തിന്റെ പേരും—11/2013-14—കമ്പ്യൂട്ടറുകളും അനുബന്ധ ഉപകരണങ്ങളും ലാപ്ടോപ്പും.

ദർഘാസ് ഫോറത്തിന്റെ വില:

അസൽ— 2,600 + 5% വാറ്റ്.

പകർപ്പ്— 1,300 + 5% വാറ്റ്.

ദർഘാസുകളിൽ ടെണ്ടർ നമ്പർ രേഖപ്പെടുത്തി അഡീഷണൽ ഡയറക്ടർ (മെഡിക്കൽ), ആരോഗ്യവകുപ്പ് ഡയറക്ടറേറ്റ്, തിരുവനന്തപുരം-695 035 എന്ന വിലാസത്തിൽ 22-1-2014 പകൽ 3 മണിക്ക് മുമ്പായി സമർപ്പിക്കേണ്ടതാണ്. സ്വീകാര്യമായ ദർഘാസുകൾ നൽകിയവരുടെയോ അവർ അധികാരപ്പെടുത്തിയവരുടെയോ സാന്നിധ്യത്തിൽ 23-1-2014 പകൽ 11 മണിക്ക് തുറക്കുന്നതാണ്. ദർഘാസുകൾ വിൽക്കുന്ന അവസാന തീയതി 21-1-2014 പകൽ 3 മണിവരെയാണ്. ദർഘാസ് ഫോറം ആവശ്യമുള്ളവർ അതിന്റെ വില

മണിയോർഡർ ആയോ നേരിട്ടോ അടച്ച് തിരുവനന്തപുരം ഗവൺമെന്റ് മെഡിക്കൽ സ്റ്റോർ ആഫീസർക്ക് പ്രവൃത്തി ദിവസങ്ങളിൽ രേഖാമൂലം അപേക്ഷിച്ചാൽ ലഭിക്കുന്നതാണ്. ദർഘാസ് സമർപ്പിക്കേണ്ട അവസാന തീയതിക്ക് 10 പ്രവൃത്തി ദിവസം മുമ്പെങ്കിലും ലഭിക്കുന്ന മണിയോർഡറുകൾ മാത്രമേ സ്വീകരിക്കുകയുള്ളൂ. ദർഘാസ് ഫോറം കൈമാറ്റം ചെയ്യാൻ പാടില്ല. ദർഘാസ് ഫോറത്തിന്റെ വില ചെക്കുകളായോ ഡ്രാഫ്റ്റായോ പോസ്റ്റൽ ഓർഡർ ആയോ സ്വീകരിക്കുന്നതല്ല. അപേക്ഷയിൽ ദർഘാസ് നമ്പർ, സാധനത്തിന്റെ പേര് എന്നിവ നിർബന്ധമായും കാണിച്ചിരിക്കേണ്ടതാണ്. ദർഘാസിന്റെ പ്രത്യേക നിബന്ധനകൾ ദർഘാസ് ഫോറത്തോടൊപ്പം വാങ്ങേണ്ടതാണ്.

No. MSP1-79723/2013/DHS.

28th December 2013.

Sealed tenders are invited for the supply of following items for various Hospitals under the Health Services Department.

Tender No. and Name of Item—11/2013-14—Computer with peripherals and Laptop.

Cost of tender form :

Original— 2,600 + 5% VAT.

Duplicate— 1,300 + 5% VAT.

Last date and time of sale of tender form—
21-1-2014, 3 p. m.

Last date and time of receipt of tender—
22-1-2014, 3 p. m.

Date and time of opening of tender—23-1-2014,
11 a. m.

Tender forms will be obtained from the Stores Officer,
Government Medical Store, Directorate of Health Services,
Thiruvananthapuram-695 035 on payment of cash or
Money Order.

Those who want tender form by Money Order shall
send 50 (Rupees fifty only) as postal charge. Money
Order should reach this office on or before ten working
days from the last date of sale of tenders. Tender No.,
Name of the item should be mentioned in your request
for tender form. Postal Orders, Postal Stamps, Bank Draft
or Bank Cheque will not be accepted towards the cost of
tender. Sealed tender shall be submitted to the Additional
Director of Health Services (Medical), Directorate of
Health Services, Thiruvananthapuram.

Special terms and conditions of the tender shall be
obtained with the tender forms.

(Sd.)

Directorate of Health Services, Addl. Director of
Thiruvananthapuram. Health Services (Medical).

ലഘു ദർഘാസ് പരസ്യം

നമ്പർ സി.6/10990/2012.

2013 ഡിസംബർ 9.

ആലപ്പുഴ ജില്ലാ മെഡിക്കൽ ഓഫീസിന്റെ (ആരോഗ്യം)
കീഴിൽ ഫ്ലൂറൈഡ് നിർമ്മാണ പരിപാടിക്കുവേണ്ടി
ലാബ് എക്യൂപ്പ്മെന്റും റീ ഏജന്റും വിതരണം
ചെയ്യുന്നതിന് താല്പര്യമുള്ള സ്ഥാപനങ്ങളിൽ നിന്നും
മുദ്രവെച്ച കവറിൽ ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

നിരതദ്രവ്യം— 5,000.

ദർഘാസ് ഫോറത്തിന്റെ വില— 1,400 + വാറ്റ്.

ദർഘാസ് ഫോറം വില്പന തുടങ്ങുന്ന തീയതി—
16-1-2014.

ദർഘാസുകൾ വില്ക്കുന്നതും സ്വീകരിക്കുന്നതുമായ
അവസാന തീയതിയും സമയവും—28-1-2014, 4 മണി.

ദർഘാസുകൾ തുറക്കുന്ന തീയതിയും സമയവും—
30-1-2014, 3 മണി.

എക്യൂപ്പ്മെന്റുകളുടെയും റീ ഏജന്റുകളുടെയും വിശദ
വിവരങ്ങൾ ഓഫീസിൽ നിന്നും ലഭ്യമാണ്.

ദർഘാസുകൾ ജില്ലാ മെഡിക്കൽ ഓഫീസർ (ആരോഗ്യം),
ജില്ലാ മെഡിക്കൽ ഓഫീസ്, ആലപ്പുഴ-11 എന്ന
വിലാസത്തിലാണ് സമർപ്പിക്കേണ്ടത്.

ദർഘാസ് ഫോറുകൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ജില്ലാ
മെഡിക്കൽ ഓഫീസിൽ നിന്നും വിലയ്ക്ക് ലഭിക്കുന്നതാണ്.

വിശദവിവരങ്ങൾ 16-1-2014 മുതൽ 28-1-2014 വരെയുള്ള
എല്ലാ പ്രവൃത്തി ദിവസങ്ങളിലും രാവിലെ 10 മണി മുതൽ
വൈകിട്ട് 5 മണി വരെ ഓഫീസിൽ നിന്നും ലഭ്യമാണ്.

ദർഘാസ് പരസ്യം റദ്ദ് ചെയ്യുന്നതിനുള്ള അധികാരം
ജില്ലാ മെഡിക്കൽ ഓഫീസിൽ നിഷ്പിതമാണ്.

Technical Specifications for fluoride ion meter

Fluoride ion meter for Measurement of Fluoride in
Drinking Water, Urine Serum

Technical Features

1. Large colour graphic LCD display which can
show parameter(s) together or each parameter
individually.
2. Display should provide additional information of
showing readings with connections, time, date,
sample ID, User ID, Calibration information,
temperature, etc.
3. Up to 5 point calibration.
4. Meter should have Auto Read continuous display
facility.
5. ISE measurement should have Auto Blank, Linear
Point to Point & Non linear point to point.
6. Low Concentration Range Stability.
7. At least 1500 point Data logging with time and
date.
8. USB and RS232 for transfer of data to PC.
9. Software upgradation facility of equipment
through internet.

Technical Specifications

ISE	:	
Range	:	0 to 100 ppm
Resolution	:	Up to 3 significant digits
Units	:	mg/L, ppm, ppb
Power	:	Universal AC Power Adapter, 90—260 VAC 50—60 Hz

Accessories to be supplied with Equipment

1. Fluoride Combination, Epoxy Body Electrode - 2 Nos.
2. Fluoride Calibration Standard 1.0 ppm &
100ppm., 400ml each - 1 No.
3. TISAB III, at least 400 ml - 2 Nos.
4. TISAB II, at least 3 Litres - 1 No.
5. Electrode filling solution, 5×60ml in one
packet - 1 No.
6. Electrode holding Stand - 1 No.

The supplier should provide 3 year warranty for the equipment & 2 years AMC

Supplier should provide list of users of the equipments in Health Sector during the last five years.

2. Table Top Centrifuge without refrigeration

Technical Specifications :

- ◆ User friendly microprocessor based control with LED display
- ◆ Choice of high brake, low brake and coasting.
- ◆ Imbalance detector cut off.
- ◆ Safety lid interlock to prevent cover opening during centrifugation.
- ◆ Speed holding accuracy 100 rpm
- ◆ Self diagnostic for errors.
- ◆ Maximum speed: 20000 rpm
- ◆ Maximum rcf : 37560g
- ◆ Maximum tube size; 100ml
- ◆ Maximum capacity; 400ml
- ◆ 24 × 1.5 ml. Swing angle rotor with polypropylene tubes
- ◆ 8 × 15 ml. Swing angle rotor with polypropylene tubes.
- ◆ 6×50ml Swing angle rotor with polypropylene tubes.

3. Digital pH Meter

Technical Specifications:

pH Range	0-14 pH
pH Resolution	0.001pH
pH Accuracy	+ 0.02 pH
Temperature compensation	0-100 deg C
mV	+2000mV
mV Resolution	0.1mV
mV accuracy	+0.2mV

Features :

- ◆ Automatic temperature compensation
- ◆ Storage upto 50 test results
- ◆ Display of time based pH sampling data upto 200 samples
- ◆ Supply Voltage: 90-250V,+10%V AC, 50 Hz
- ◆ Operating temperature : 10-45 deg C
- ◆ Humidity: 5-80% Rh
- ◆ Automatic pH calibration: 3 points 4.0, 7.0 & 9.2 pH

- ◆ Manual pH calibration: 3 points using known buffers
- ◆ Display: Backlite LCD
- ◆ pH sensor: Glass pH electrode
- ◆ Replaceable electrode.

4. Mettler Balance

Technical Specifications:

- ◆ Electronic analytical top loading Balance (0.0001g) precision.
- ◆ Weighing range: 220g
- ◆ Calibration: Built in & automatic.
- ◆ Antitheft protection password and mechanical
- ◆ Below balance weighing.
- ◆ Net total application.
- ◆ Add up summation.
- ◆ Density determination & animal weighing.
- ◆ Readability : 0.1mg.
- ◆ Linearity : ± 0.2 mg
- ◆ Reproducibility: ≤ 0.1 mg
- ◆ Pan size: 80mm dia.
- ◆ Response time : 4 seconds.
- ◆ Display : Large LCD
- ◆ Draft shield: Open able & built in.

Features.—Piece counting, percent weighing and weighing in most common international units.

5. Magnetic Mixer

Technical Specifications:

(Magnetic stirrer)

- ◆ The stainless steel top housing series of magnetic stirrers are most compact models.
- ◆ PMDC motor gives high torque even at lower speeds and maintains speed stability despite viscosity or volume changes.
- ◆ Accurate steeples speed control allows smooth variation up to 1200 rpm.
- ◆ Different size Teflon coated magnetic bar.

Salient Features:

- ◆ PMDC motor for higher torque even at low speeds
- ◆ Better low speed stirring even with small volume
- ◆ Accurate step less speed control maintains excellent speed stability.

6. Incubator

Technical Specifications:

- Incubators should be sturdy, with double walled construction with complete inner chamber made of Highly Polished Stainless Steel.
- Outer chamber should be made of Mild Steel Sheet, finished with powder coated paint, 75mm gap between the walls should be filled with special grade glass wool for proper insulation and to avoid heat losses.
- Inner chamber should be fabricated with ribs for adjusting shelves to convenient height.
- Should be supplied with 2 or 3 removable shelves. SHELVES should be made of fully Polished Stainless Steel Sheet as per chamber.
- Door should be insulated & fitted with heavy hinges.
- Door should have Double glass window which facilitate inspection of samples without opening the door.
- Heating elements should be made of high grade imported nichrome wire and put inside the porcelain beads and placed at the bottom and side ribs for uniform temperature all over the space.
- Temperature range should be 25°C-60°C
- Temperature Control:—Temperature control should be microprocessor based with PID Controller.
- Air circulation should be provided.
- Ventilation:—Air ventilators should be provided on both sides at top flue away any gas or fumes.
- Control panel:— The equipment should be provided with a panel having a thermostat control knob, on/off switch and two pilot lamps.
- Power Requirement:—Supplied with cord and plug suitable to operate on 220V single phase, 50 Hz, AC supply.
- Chamber Size: 45×45×45 cms.

7. Micropipettes

Technical Specifications:

- Rugged construction.
- Extremely smooth piston movement.
- Fully Autoclavable.
- Easy to maintain.
- Individual Factory Calibration Certificate with each pipette.
- Volume range 5 to 50 µl
- Volume range 100 to 1000 µl

ജില്ലാ മെഡിക്കൽ ഓഫീസ്
(ആരോഗ്യം), ആലപ്പുഴ.

(ഒപ്പ്)
ജില്ലാ മെഡിക്കൽ ഓഫീസർ
(ആരോഗ്യം).

ടെണ്ടർ നോട്ടീസ്

നമ്പർ ബി-1085/2013.

2013 ഡിസംബർ 11.

കൊരട്ടി ഗവൺമെന്റ് ലെപ്രസി ആശുപത്രിയിലെ അന്തേവാസികളുടെ ആവശ്യത്തിനായി താഴെ പറയുന്നവ വിതരണം ചെയ്യുന്നതിനുള്ള മുദ്രവച്ച ടെണ്ടറുകൾ ക്ഷണിക്കുന്നു.

പ്രവൃത്തിയുടെ വിവരം—മൈക്രോ സെല്ലുലർ റബ്ബർ ചപ്പുകൾ—300 ജോഡി (ആശുപത്രി വാർഡുകൾ സന്ദർശിച്ച് അന്തേവാസികളുടെ കാലിന്റെ അളവ് ആവശ്യാനുസരണം എടുത്ത്) നിശ്ചിത മാനദണ്ഡമനുസരിച്ചുള്ള ചെറുപ്പുകൾ നിർമ്മിച്ച് നൽകണം.

ടെണ്ടർ ഫോറങ്ങൾ ആശുപത്രി ഓഫീസ് പ്രവൃത്തി സമയങ്ങളിൽ നേരിട്ട് തുക അടച്ച് കൈപ്പറ്റാവുന്നതാണ്.

450 + 5% വാറ്റ് ആണ് ഒരു ഫോറത്തിന്റെ വില. ഈ തുക ഒരു കാരണവശാലും മടക്കി നൽകുന്നതല്ല. ടെണ്ടർ ഫോറങ്ങൾ കൈമാറ്റം ചെയ്യുവാൻ പാടില്ല. ടെണ്ടർ ഫോറങ്ങളുടെ വിൽപന 29-1-2014-ന് ഉച്ചയ്ക്ക് 3 മണിക്ക് അവസാനിക്കുന്നതാണ്.

സുപ്രണ്ട്, ഗവൺമെന്റ് ലെപ്രസി ആശുപത്രി, കൊരട്ടിയുടെ പേരിൽ പുരിപ്പിച്ച ടെണ്ടറുകൾ 30-1-2014 (വ്യാഴാഴ്ച) ഉച്ചയ്ക്ക് 12 മണിവരെ ഈ ആശുപത്രിയുടെ ഓഫീസിൽ സ്വീകരിക്കുന്നതും, അന്നേ ദിവസം 12.30-ന് ഹാജരാകുന്നവരുടെ സാന്നിധ്യത്തിൽ തുറന്ന് പരിശോധിക്കുന്നതുമാണ്. ടെണ്ടറിൽ രേഖപ്പെടുത്തുന്ന തുകയുടെ 1% നിരതദ്രവ്യം ഡി. ഡി. ആയി സുപ്രണ്ടിന്റെ പേരിൽ ടെണ്ടറിനോടൊപ്പം സമർപ്പിക്കേണ്ടതാണ്.

ടെണ്ടർ സംബന്ധമായ എല്ലാ തീരുമാനങ്ങളും സ്ഥാപന സുപ്രണ്ടിൽ നിക്ഷിപ്തമായിരിക്കും.

സാധനങ്ങളുടെ ഗുണനിലവാരം പർച്ചേസ് കമ്മിറ്റിയുടെ പരിശോധനയ്ക്ക് വിധേയമായിരിക്കും.

കൂടുതൽ വിവരങ്ങൾക്ക് പ്രവൃത്തി ദിവസങ്ങളിൽ ഓഫീസുമായി ബന്ധപ്പെടേണ്ടതാണ്.

ഗവ. ലെപ്രസി ആശുപത്രി,
കൊരട്ടി.

(ഒപ്പ്)
സുപ്രണ്ട്.

ദർഘാസ് പരസ്യം

നമ്പർ സി2-691/2013-14.

2013 നവംബർ 18.

കോഴിക്കോട് ചേവായൂർ സർക്കാർ കുഷ്ഠരോഗ ആശുപത്രിയിലേക്ക് 300 എം. സി ആർ. ചപ്പൽ അളവെടുത്ത് ഉണ്ടാക്കി നൽകുന്നതിന് നിശ്ചിത ഫോറത്തിൽ സീൽ ചെയ്ത ടെണ്ടറുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

ടെണ്ടർ അയയ്ക്കുന്ന കവറിനുമുകളിൽ നമ്പർ സി2/691/2013-14 തീയതി 18-11-2013 എന്ന് കാണിച്ചിരിക്കണം.

ടെണ്ടർ ഫോറത്തിന്റെ വില (ടാക്സടക്കം):

ഒറിജിനൽ— 450.

ഡ്യൂപ്ലിക്കേറ്റ്— 225.

ടെണ്ടർ ഫോറം വിൽക്കുന്ന അവസാന തീയതിയും സമയവും —18-1-2014, 3 മണി.

ടെണ്ടർ കിട്ടേണ്ട അവസാന തീയതിയും സമയവും—20-1-2014, 3 മണി.

ടെണ്ടർ തുറക്കുന്ന തീയതിയും സമയവും—24-1-2014, 2 മണി.

ടെണ്ടറിൽ പറയുന്ന നിരക്കുകൾ പ്രാബല്യത്തിലിരിക്കേണ്ട തീയതി—30-6-2014.

ടെണ്ടർ ഫോറം കിട്ടുന്നതും പൂരിപ്പിച്ച് അയയ്ക്കേണ്ട തുമായ വിലാസം—സുപ്രണ്ട്, സർക്കാർ കുഷ്ഠരോഗാശുപത്രി, ചേവായൂർ, കോഴിക്കോട്-673 017.

No. C2-691/2013-14. 18th November 2013.

Sealed tenders in the prescribed forms are invited for the supply of MCR Chappels—300 Pairs to Govt. Leprosy Hospital Chevayur, Kozhikode.

Superscription tender number should be specifically noted as C2-691/2013-14 dated 18-11-2013.

Price of tender form (including Taxes) :

Original— 450.

Duplicate— 225.

(ST Cash or Treasury Chalan)

Date and time of close of sale of tender form— 18-1-2014 at 3 p.m.

Date of receipt of tender—20-1-2014.

Date and time for opening of tender—24-1-2014, at 2 p. m.

Date upto which rates are to be firms—30-6-2014.

Address of officer from whom tender forms are to be obtained and whom tenders are to be sent—The Superintendent, Govt. Leprosy Hospital, Chevayur, Kozhikode- 673 017.

Govt. Leprosy Hospital,
Chevayur, Kozhikode.

(Sd.)
Superintendent.

ദർഘാസ് പരസ്യം

നമ്പർ ബി2-20816/2013. 2013 ഡിസംബർ 28.

ദേശീയ അന്ധതാ നിവാരണ പരിപാടിയുടെ ഭാഗമായി കണ്ണൂർ ജില്ലയിലെ സ്കൂൾ കുട്ടികൾക്കായി ഏകദേശം 1300 കണ്ണടകൾ വിതരണം ചെയ്യുന്നതിന് താല്പര്യമുള്ള സ്ഥാപനങ്ങളിൽനിന്നും അംഗീകൃത ഡീലർമാരിൽനിന്നും മൽസര സ്വഭാവമുള്ള ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

ഇനവിവരം

കണ്ണടകൾ കവർ/പെട്ടിയോട് കൂടിയായിരിക്കണം

സി. ആർ. ലെൻസ് ആയിരിക്കണം

ലോഹനിർമ്മിത ഫ്രെയിം-ഗ്യാരണ്ടിയോടുകൂടി കറുപ്പ്, ബ്രൗൺ, സിൽവർ എന്നീ നിറങ്ങളിൽ

സൈഡ് സ്ക്വിങ് ആയിരിക്കണം.

(കണ്ണടകൾ വിവിധ സ്കൂളുകളിൽ ആവശ്യാനുസരണം വിതരണം നടത്തേണ്ടതാണ്.)

ദർഘാസുകൾ വിതരണം ചെയ്യുന്നതും സ്വീകരിക്കുന്ന തുമായ സ്ഥലം—ജില്ലാ മെഡിക്കൽ ഓഫീസ് (ആരോഗ്യം), സിവിൽ സ്റ്റേഷൻ പി. ഒ., കണ്ണൂർ.

അടങ്കൽ തുക — 3,58,000

നിരതദ്രവ്യം—അടങ്കൽ തുകയുടെ 1% (ഡിമാന്റ് ഡ്രാഫ്റ്റ്).

ദർഘാസ് ഫോറത്തിന്റെ വില :

അസ്സൽ— 800 + വാറ്റ്.

പകർപ്പ്— 400 + വാറ്റ്.

ടെണ്ടർ ഫോറം വിതരണം ചെയ്യുന്ന അവസാന തീയതി—20-1-2014 അപരാഹ്നം.

ടെണ്ടർ ഫോറങ്ങൾ സ്വീകരിക്കുന്ന അവസാന തീയതി— 21-1-2014 അപരാഹ്നം.

ടെണ്ടർ തുറക്കുന്ന തീയതി—22-1-2014 അപരാഹ്നം.

ദർഘാസിനോടൊപ്പം 100-ന്റെ മുദ്രക്കടലാസിൽ നിർദ്ദിഷ്ട മാതൃകയിൽ എഴുതി ഒപ്പിട്ട പ്രാഥമിക കരാർ ഉടമ്പടിയും ഉള്ളടക്കം ചെയ്യേണ്ടതാണ്. ദർഘാസ് സംബന്ധിച്ച് നിലവിലുള്ള എല്ലാ നിബന്ധനകളും ഇതിനും ബാധകമായിരിക്കും. ദർഘാസ് കിട്ടിയാൽ ദർഘാസ് തുകയുടെ 5% സെക്യൂരിറ്റി തുകയായി ജില്ലാ മെഡിക്കൽ ഓഫീസറുടെ (ആരോഗ്യം) പേരിൽ ഡിമാന്റ് ഡ്രാഫ്റ്റായി കെട്ടിവേക്കേണ്ടതാണ്. ദർഘാസ് ലഭിച്ചവർ വിതരണ ഉത്തരവ് ലഭിക്കുന്ന മുറയ്ക്ക് ഉടൻ തന്നെ വിതരണം ചെയ്യേണ്ടതാണ്. ദർഘാസ് സംബന്ധിച്ച കൂടുതൽ വിവരങ്ങൾ കണ്ണൂർ ജില്ലാ മെഡിക്കൽ ഓഫീസിൽ (ആരോഗ്യം) നിന്നും പ്രവർത്തി സമയങ്ങളിൽ അറിയാവുന്നതാണ്.

(ഒപ്പ്)
ജില്ലാ മെഡിക്കൽ ഓഫീസ്, ജില്ലാ മെഡിക്കൽ ഓഫീസർ
കണ്ണൂർ. (ആരോഗ്യം).